


28-9909-45 Superdex™ 200 Increase 5/150 GL

28-9909-44 Superdex 200 Increase 10/300 GL



Please read these instructions carefully before using the columns.

Intended use

Superdex 200 Increase 5/150 GL and Superdex 200 Increase 10/300 GL columns are intended for research use only, and shall not be used in any clinical or *in vitro* procedures for diagnostic purposes.

Safety

For use and handling of the products in a safe way, please refer to the Safety Data Sheet.

Quick information

Superdex 200 Increase 5/150 GL and Superdex 200 Increase 10/300 GL are pre-packed Tricorn™ glass columns.

Superdex 200 Increase 5/150 GL is for rapid size analysis of proteins and other biomolecules. Short cycle times, together with small sample volume and low consumption of buffer make this column ideal for use in screening experiments to check protein homogeneity.

Superdex 200 Increase 10/300 GL is suitable for small scale preparative purification (µg-mg) as a final polishing step, as well as for protein analysis and characterization.

The columns are supplied with two fingertight connectors 1/16" male for connection to ÄKTA™ or other systems.

Table 1. Medium and column data

Medium data			
Matrix	Composite of cross-linked agarose and dextran		
Average Particle size	8.6 µm		
Exclusion limit (M _r)	Approx. 1.3 × 10 ⁶		
Optimum separation range			
globular proteins (M _r)	10 000 to 600 000		
dextran (M _r)	1000 to 100 000		
pH stability range			
regular use	3 to 12		
cleaning	1 to 14		
Temperature			
operating	4°C to 40°C		
storage	4°C to 30°C		
Column data		5/150	10/300
Bed dimensions (mm)	5 × 153-158	10 × 300-310	
Approximate bed volume (ml)	3	24	
Column efficiency (N/m)	> 42 000	> 48 000	
Flow rate ¹			
recommended (ml/min)	0.45	0.75	
maximum (ml/min)	0.75	1.80	
Pressure drop over packed bed ²	< 5.0 MPa, 50 bar, 725 psi	< 4.0 MPa, 40 bar, 580 psi	
Column hardware pressure limit	10 MPa, 100 bar, 1450 psi	5.0 MPa, 50 bar, 725 psi	

¹ Water at room temperature.

² Determine the limit according to section *Setting column pressure limits*.



First time use




Fig 1. Illustration of how to lock the adapter. The locking ring (black) must be in the down-position to prevent uncontrolled adjustment of the column's bed height.

Before connecting the column to a chromatography system, make sure there is no air in the tubing and valves. Remove the storage/shipping device, see section *Delivery/storage*, and the stop plug from the column. Check that the upper adapter is locked (locking ring pressed down, see Fig. 1). Make sure that the column inlet is filled with liquid and connect it drop-to-drop to the system. For maximum resolution especially on the 5/150 column, minimize dead volumes between the injection valve and the column as well as between the column outlet and the detector.

Prepare the column for first-time use as follows:

- Equilibrate with at least 2 column volumes (CV) of room tempered water at a flow rate of 0.30 ml/min (5/150) or 0.75 ml/min (10/300).
- Set pressure limits for the column in your method according to section *Setting column pressure limits*.
- Equilibrate with at least 2 CV eluent at a flow rate of 0.30 ml/min (5/150) or 0.75 ml/min (10/300).
- It is recommended to perform a column performance control for future comparisons, see section *Column performance control*.



NOTICE

Make sure to not exceed the pressure limits of the column. This is particularly important when working at low temperatures, like in a cold room, or when the column is used with 20% ethanol or other viscous solutions. Set pressure limits according to section *Setting column pressure limits*.

Setting column pressure limits

There are two pressure limits to consider when running the column, the pressure drop over the packed bed and the column hardware pressure limit. The pressure drop over the packed bed differ for each column and the limit has to be individually set. Note that it might be significantly lower compared to the maximum value noted in Table 1.

Exceeding any pressure limit may lead to collapse of the gel bed or damage to the column hardware. Increased pressure is for example generated when running/using one or a combination of the following parameters:

- Eluent or sample with high viscosity compared to water
- Low temperature compared to room temperature
- Modifications to the flow path, for example changing to thinner/longer tubing

For optimal functionality it is important to know the pressure drops over different parts of your system and how they affect the column. All ÄKTA chromatography systems measure pressure at the system pump, p_{pump} (see Fig. 2). Some systems have additional pressure sensors located before and after the column, $p_{\text{pre-cp}}$ and $p_{\text{post-cp}}$.




Fig 2. Example of the pressure in different parts of a system during run of a column.

- Δp_{before} does not affect the column.
- The pressure on the column hardware is the sum of Δp_{after} and Δp . Do not exceed the column hardware limit!
- Δp , is individual for each column and needs to be determined.

For more information, please refer to the *ÄKTA laboratory-scale Chromatography Systems Instrument Management Handbook*.

How to set pressure limit for ÄKTAexplorer™, ÄKTApurifier™, ÄKTAFPLC™, ÄKTAmicro™, ÄKTA pure (Column Valve V9-Cs) and other systems with a pressure sensor in the pump

Determination of column specific pressure drop over the packed bed, Δp (see Fig. 2):

- 1 Δp_{before} is measured in absence of the column. Run the pump at maximum flow rate of the column (see Table 1) with water at room temperature. Let the flow drip from the tubing that will later be connected to the column. Note the pressure as Δp_{before} .
- 2 Connect the inlet of the, with room tempered water equilibrated, column to the system. Run the pump at the same flow rate as in step 1, let the flow drip from the column outlet. Note the pressure value.
- 3 Δp is calculated as the pressure value in step 2 minus Δp_{before} . The Δp value will be used in step 7 below.

Setting pressure limit in method at your experimental conditions (intended system setup, flow rate, temperature and eluent):

- 4 Δp_{before} is measured in absence of the column. Run the pump at your intended flow rate. Let the flow drip from the tubing that will later be connected to the column. Note the pressure as Δp_{before} .
- 5 Instead of the column, connect a piece of tubing¹ to the system. Run the pump at the same conditions as in step 4. Note the pressure value as the total system pressure.
- 6 Δp_{after} is calculated as the total system pressure value noted in step 5 minus Δp_{before} , noted in step 4.
- 7 Calculate $\Delta p + \Delta p_{\text{after}}$. If this value is lower than the column hardware pressure limit (see Table 1), set the pressure limit in your method as $\Delta p + \Delta p_{\text{after}} + \Delta p_{\text{before}}$.

You can now start your experiment!

If $\Delta p + \Delta p_{\text{after}}$ exceed the column hardware pressure limit, reduce the flow rate or Δp_{after} . Repeat step 4-7.

How to set pressure limit for ÄKTA avant and ÄKTA pure (Column Valve V9-C)

Δp and $p_{\text{pre-cp}}$ (see Fig. 2) are automatically monitored by the system. Note that the measured values includes the tubing used to connect the column to the instrument.

Setting pressure limit in method:

- 1 Check that the **max pre column pressure alarm** is set to the same as the column hardware pressure limit (see Table 1).
- 2 Connect the column (equilibrated with room tempered water) to the system. Start running the column at a low flow rate (e.g., half recommended flow rate) with water/eluent at room temperature. Slowly increase the flow until the maximum flow rate of the column or the limit set in step 1 is reached (see Table 1). Note the pressure over the packed bed, Δp , (Delta C pressure) and set the value as **max delta column pressure**.

You can now start your experiment!

Column performance control

In order to detect any changes in column performance, it is very important that you make an initial test with your particular system configuration. Note that the contribution from dead volumes in the instrument to band broadening will vary depending on system set-up and will influence column efficiency, thus the obtained efficiency on your system might be lower compared to the specifications in Table 1.

Column efficiency test

Column efficiency, expressed as the number of theoretical plates per meter, N/m , is calculated using the following equation:

$$N/m = 5.54 \times (V_R / W_h)^2 \times (1000/L)$$

where

N/m	=	number of theoretical plates/meter
V_R	=	volume eluted from the start of sample application to the peak maximum
W_h	=	peak width measured as the width of the recorded peak at half of the peak height
L	=	bed height (m)

Column Superdex 200 Increase 5/ 150 GL

Check the performance of the column using the following procedure:

Sample:	10 μ l 2% acetone (20 mg/ml) in buffer or water
Eluent:	Buffer or water
Flow rate:	0.30 ml/min, room temperature
Detection:	280 nm

Column Superdex 200 Increase 10/300 GL

Check the performance of the column using the following procedure:

Sample:	100 μ l 2% acetone (20 mg/ml) in buffer or water
Eluent:	Buffer or water
Flow rate:	1.0 ml/min, room temperature
Detection:	280 nm

Function test

As an alternative to the above efficiency test, check the column performance by running a function test.

Column Superdex 200 Increase 5/ 150 GL.

Sample:	1. Thyroglobulin (M_r 669 000) 3 mg/ml
	2. Aldolase (M_r 158 000) 3 mg/ml
	3. Conalbumin (M_r 75 000) 3 mg/ml
	4. Carbonic anhydrase (M_r 29 000) 3 mg/ml
	5. Ribonuclease A (M_r 13 700) 3 mg/ml

Sample volume:	10 μ l
Eluent:	0.01 M phosphate buffer, 0.14 M NaCl, pH 7.4
Flow rate:	0.45 ml/min, room temperature
Detection:	280 nm

Result is shown in Fig. 3.

¹ Avoid thin and/or long tubing that will give back pressure.




Fig 3. Typical chromatogram from a function test of Superdex 200 Increase 5/150 GL using ÄKTAmicro.

Column Superdex 200 Increase 10/300 GL

Sample:	1. Thyroglobulin (M_r 669 000) 3 mg/ml
	2. Ferritin (M_r 440 000) 0.3 mg/ml
	3. Aldolase (M_r 158 000) 3 mg/ml
	4. Conalbumin (M_r 75 000) 3 mg/ml
	5. Ovalbumin (M_r 44 000) 3 mg/ml
	6. Carbonic anhydrase (M_r 29 000) 3 mg/ml
	7. Ribonuclease A (M_r 13 700) 3 mg/ml

Sample volume: 100 μ l
 Eluent: 0.01 M phosphate buffer, 0.14 M NaCl, pH 7.4
 Flow rate: 0.75 ml/min, room temperature
 Detection: 280 nm

Result is shown in Fig. 4.




Fig 4. Typical chromatogram from a function test of Superdex 200 Increase 10/300 GL using ÄKTApurifier.

Try these conditions first

Eluent:	0.01 M phosphate buffer, 0.14 M NaCl, pH 7.4
Flow rate:	0.45 ml/min (5/150); 0.75 ml/min (10/300)
Sample volume:	10 μ l (5/150); 100 μ l (10/300)

Equilibration is not necessary between runs with the same buffer. Read the section *Optimization* for information on how to optimize a separation.

Sample recommendations

Molecular weight (M_r):	10 000 to 600 000
Protein concentration:	Up to 50 mg/ml in sample, for higher resolution below 10 mg/ml.
Sample volume:	4 to 50 μ l (5/150) 25 to 500 μ l (10/300)
Preparation:	Dissolve the sample in eluent, filter through a 0.22 μ m filter or centrifuge at 10 000 g for 10 min.

System recommendations

The small bed volume of the 5/150 column makes it sensitive to dead volumes in the system. For this column it is recommended to use systems like ÄKTApurifier 10, ÄKTA pure, ÄKTAmicro or similar. Use short, narrow capillaries and avoid unnecessary components in the flow path.

Note: *Be aware of pressure limits.*

The 10/300 column has a larger bed volume and can be used in systems like ÄKTAexplorer 10 and ÄKTA avant 25 in addition to the systems mentioned above.

Chromatogram 5/150 column on different systems




Fig 5. Comparison of protein separation on Superdex 200 Increase 5/150 on different systems. Due to different UV cells for the systems, mAU scales differ.

Chromatogram 10/300 column on different systems




Fig 6. Comparison of protein separation on Superdex 200 Increase 10/300 on different systems. Due to different UV cells for the systems, mAU scales differ.

Delivery/storage

The column is delivered with a storage/shipping device that prevents it from drying out. The column is equilibrated with degassed 20% ethanol.

If the column is to be stored for more than 2 days after use, wash the column with 2 column volumes (CV) of water and then equilibrate with at least 2 CV 20% ethanol.

Note: Use a lower flow rate for 20% ethanol.

We recommend that you connect the storage/shipping device according to section *How to connect the storage/shipping device* for long term storage.



How to remove the storage/shipping device

- 1 Push down the spring-loaded cap.
- 2 Remove the locking pin.
- 3 Release the cap and unscrew the device.



How to refill the storage/shipping device

- 1 Connect a syringe or pump to the storage/shipping device and fill with 20% ethanol over the mark on the tube. Remove the syringe or connection to the pump.
- 2 Tap out air bubbles and push the plunger to the mark on the device.



How to connect the storage/shipping device

- 1 Fill the column inlet and luer connector with 20% ethanol and connect the filled storage/shipping device drop-to-drop to the top of the column.
- 2 Mount the spring-loaded cap (2) and secure it with the locking pin (3).

The glass tube is coated with a protecting plastic film. Small quantities of air may occasionally be trapped between the glass and the film during manufacture. The resulting uneven surface does not affect column performance or durability.

Choice of eluent

Select an eluent that ensures the sample is fully soluble. Also try to choose an eluent that will simplify downstream applications. For example, if the proteins/peptides are to be lyophilized, a volatile eluent is necessary. Since ionic interactions can occur with both acidic and basic proteins at very low salt concentrations, a recommended buffer is 0.01 to 0.05 M sodium phosphate with additional 0.15 M NaCl, pH 7.4. Table 2 lists some useful eluent compositions.

Table 2. Useful eluent compositions

pH	Buffer/eluent	Properties/application examples
5.0	0.1 M ammonium acetate	Good solubility for some enzymes, e.g., cellulases. Volatile.
6.8	0.2 M sodium phosphate	Suitable for some antibody separations.
7.2	0.05 M phosphate + 0.15 M NaCl	Physiological conditions.
7.8	0.15 M ammonium hydrogen carbonate	Suitable for some DNA and protein separations. Volatile. Should be used fresh.
8.0	0.1 M Tris-HCl, 0.001 M EDTA	Very good solubility for DNA and RNA.
8.6	6 M guanidine hydrochloride in 0.05 M Tris-HCl	Good UV-transparency. Suitable if there is a need to purify proteins under denaturing conditions.
11.5	0.05 M NaOH	Good solubility for some compounds.

Buffer additives

Up to 8 M urea (pH<7)	Good solubility for many components. Biological activity can be maintained at lower urea contents. Certain risk for carbamylation of proteins.
6 M guanidine hydrochloride	Molecular weight determinations of subunits.
0.1% SDS, Tween or similar	Good solubility for some proteins, e.g., membrane proteins. Make sure you equilibrate completely with the detergent solution.
0.2 M arginine	Decrease tendency of aggregation.

Buffers and solvent resistance

De-gas and filter all solutions through a 0.22 μm filter. Install an on-line filter before the injection valve.

Note: Buffers and solvents with increased viscosity will affect the back pressure. Reduce the flow rate if necessary.

Long term use

Long term use refers to use where the medium is stable over a long period of time without adverse side effects on its chromatographic performance.

- All commonly used aqueous buffers, pH 3 to 12
- Urea, up to 8 M
- Ionic and non-ionic detergents
- Guanidine hydrochloride, up to 6 M

Short term use

Short term use refers to the use during regeneration, cleaning-in-place, and sanitization procedures.

- Acetonitrile, up to 30%
- Sodium hydroxide, up to 1 M
- Ethanol, up to 70%
- Methanol, up to 100%
- Acetic acid, up to 1 M
- Isopropanol, up to 30%
- Hydrochloric acid, up to 0.1 M
- Trifluoroacetic acid, up to 10%
- Formic acid, up to 70%

Avoid:

- Oxidizing agents
- Unfiltered solutions

Optimization

If your results are unsatisfactory, consider the following actions.

Flow rate

Action: Decrease the flow rate.

Effect: Improves resolution for high molecular weight components. The resolution for small components may be decreased.




Fig 7. Comparison of protein separation on Superdex 200 Increase 5/150 at different flow rates.




Fig 8. Comparison of protein separation on Superdex 200 Increase 10/300 at different flow rates.

Sample volume

Action: Decrease the sample volume.

Effect: Improves resolution.




Fig 9. Comparison of protein separation on Superdex 200 Increase 5/150 using different sample volumes.




Fig 10. Comparison of protein separation on Superdex 200 Increase 10/300 using different sample volumes.

System dead volumes

Action: Decrease system dead volumes.

Effect: Improves resolution.




Fig 11. Comparison of protein separation on Superdex 200 Increase 5/150 using different diameters of a 32 cm long capillary connected to the column.

Organic solvent

Action: Add organic solvent.

Effect: Increases solubility of hydrophobic compounds. Changes selectivity.

Columns in series (Superdex 200 Increase 5/150)

Action: Connect two columns in series.

Effect: Increases resolution due to increased bed height. Back pressure will increase. Be sure not to exceed maximum pressure limits.




Fig 12. Comparison of protein separation on one or two Superdex 200 Increase 5/150 connected in series.

For more information, please refer to the handbook *Gel filtration, Principles & Methods*.

Cleaning-in-place (CIP)

Perform the following regular cleaning cycle after 10 to 20 separation cycles.

Note: When performing CIP, reversed flow is recommended.

Regular cleaning:

- 1 Wash the column with 1 column volume (CV) of 0.5 M sodium hydroxide alternatively 0.5 M acetic acid at a flow rate of 0.13 ml/min (5/150) or 0.5 ml/min (10/300).
- 2 Immediately rinse the column with 1 CV water followed by at least 2 CV eluent at a flow rate of 0.13 ml/min (5/150) or 0.5 ml/min (10/300).

Before the next run, equilibrate the column until the UV baseline and pH are stable. Check the column according to section *Column performance control* that the performance is restored.

More rigorous cleaning:

- Depending on the nature of the contaminants, one of the cleaning solutions in the section *Buffers and solvent resistance* may be used. Always rinse with at least 2 CV water after any of the cleaning solution has been used.
- If column performance is not restored, wash the column with 3 CV 0.5 M arginine. Rinse with at least 2 CV water.
- If column performance is still not restored, inject a solution of 1 mg/ml pepsin in 0.1 M acetic acid containing 0.5 M NaCl and leave overnight at room temperature or one hour at 37°C. After enzymatic treatment, clean the column according to the procedure described in the section *Regular cleaning*.
- Change the filter at the top of the column. (Since contaminants are introduced with the liquid flow, many of them are caught by the filter). Instructions for changing the filter are supplied with the Filter Kit. Perform a regular cleaning as described above.
- If necessary, suspend 2 to 3 mm of the top of the gel bed and remove it with a Pasteur pipette. Adjust the adaptor to eliminate the space above the gel.

Troubleshooting

Symptom	Remedy
Increased back-pressure over the column and/or loss of resolution.	Confirm that the column is the cause (see below). If so, clean it according to the procedure described in the section <i>Cleaning in place (CIP)</i> . To confirm that the high back-pressure in the system is caused by the column, disconnect one piece of equipment at a time (starting at the fraction collector) with the pumps running. Check the pressure reading after each piece has been disconnected to determine the source of the back-pressure.
Air in the column	Note that small amounts of air will normally not affect the performance of the column. Run 3 to 4 column volumes (CV) of well de-gassed eluent in an <u>upflow direction</u> at a flow rate of 0.2 ml/min (5/150) or 0.5 ml/min (10/300) at room temperature.
Space between gel bed and adapter	Turn down the adapter to the gel bed. Perform a Column performance control.

Ordering information

Product	No. per pack.	Code No.
Superdex 200 Increase 5/150 GL	1	28-9909-45
Superdex 200 Increase 10/300 GL	1	28-9909-44

Related products

Product	No. per pack.	Code No.
Superdex 200 Increase 3.2/300	1	28-9909-46
Gel filtration LMW Calibration Kit	1	28-4038-41
Gel filtration HMW Calibration Kit	1	28-4038-42

Accessories

Product	No. per pack.	Code No.
Tricorn 10 Filter Kit	1	18-1153-11
Tricorn 5 Filter kit	1	18-1153-02
Filter tool	1	18-1153-20
Fingertight connector, 1/16" male	10	18-1112-55
Tricorn storage/shipping device	1	18-1176-43

Literature

Handbook	Code No.
Gel filtration Principles & Methods	18-1022-18
ÄKTA laboratory-scale Chromatography Systems Instrument Management Handbook	29-0108-31

For local office contact information, visit www.gelifesciences.com/contact

GE Healthcare Bio-Sciences AB
Björkgatan 30
751 84 Uppsala
Sweden

www.gelifesciences.com

GE, imagination at work and GE monogram are trademarks of General Electric Company.

ÄKTA, ÄKTAexplorer, ÄKTAFPLC, ÄKTAmicro, ÄKTApurifier, Superdex and Tricorn are trademarks of GE Healthcare companies.

©2013 General Electric Company—All rights reserved.
First published Jan 2013.

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

GE Healthcare UK Ltd
Amersham Place, Little Chalfont, Buckinghamshire, HP7 9NA, UK

GE Healthcare Bio-Sciences Corp
800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA

GE Healthcare Europe GmbH
Munzinger Strasse 5, D-79111 Freiburg, Germany

GE Healthcare Japan Corporation
Sanken Bldg. 3-25-1, Hyakunincho, Shinjuku-ku, Tokyo 169-0073, Japan



imagination at work